

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	101295
Date Issued	DEC2023
Expiration Date	30JUN2024

Population	% of Population	Results (Pass/Fail)
Lymphocyte Mimics	28-36%	Pass
Monocyte Mimics	28-36%	Pass
Granulocyte Mimics	28-36%	Pass

*"% of Population" is measured by FSC and SSC

Jonathan Yang

12 / 08 / 2023

Approved by: _____

Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: 11-0538v1