

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	101197
Date Issued	NOV2023
Expiration Date	31OCT2024

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes

Section B: MFI*

Population	MFI specification	Results	Pass/Fail
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	$\geq 400,000$	2,105,960	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	$\geq 400,000$	604,766	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	$\geq 400,000$	787,896	Pass

*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by: Jonathan Yang Director, Quality Assurance 11 / 10 / 2023
Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0530v2