

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product LOT</b>	102460
<b>Date Issued</b>	MAY2025
<b>Date of Expiration*</b>	MAY2026

**\*Product has a shelf life of six months from the date of thaw. Follow the Date of Expiration if it occurs before shelf life expiration.**


## Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)
7AAD	Yes
Ghost	Yes

## Section B: MFI\*\*

Population	MFI specification	Results (Pass/Fail)
Positive for 7AAD	≥ 20,000	Pass
Positive for Ghost	≥ 200,000	Pass

**\*\*when analyzed using on Cytex Aurora 5 laser Cytometer under recommended gain settings**

**Approved by:**  **Jonathan Yang**, Director, Quality and Regulatory      5/14/2025  
Signed by: B81B0F0EDCC0144B...  
 Quality Assurance