

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102622
Date Issued	AUG2025
Expiration Date	FEB2027


Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone: SK1)	≥ 400,000	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC; clone: PC61)	≥ 400,000	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone: 145-2C11)	≥ 400,000	Pass

*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by:  Signed by: **Simon Au** Quality Engineer 9/19/2025
F2D266B88A494B3...

Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0601