

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only (RUO). Not for use in diagnostic procedures.
Product Lot	102560
Date Issued	JUL2025
Expiration Date	JUL2027

Population	MFI Specification	Results
SpectraComp eGFP Negative	< 3000	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
SpectraComp eGFP Low	10,000-30,000	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
SpectraComp eGFP Medium	100,000-250,000	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
SpectraComp eGFP High	1M-2.5M	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail

***when analyzed using Cytex Aurora 5 Laser Cytometer under recommended gain settings**

Approved by: Signed by:
Camille Shah 2/12/2026 Senior QA Analyst
D62A2FF04F7F446...
Quality Assurance