

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102510
Date Issued	JUN2025
Expiration Date	JUN2026


Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes
Rabbit	Yes
Human	Yes

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti-human CD3 FITC clone-OKT3)	$\geq 400,000$	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	$\geq 400,000$	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	$\geq 400,000$	Pass
Rabbit Positive (stained with Rabbit anti-human kappa FITC)	$\geq 400,000$	Pass
Human Positive (stained with Human IgG1 FITC clone QA16A12)	$\geq 400,000$	Pass

*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by:  Quality Engineer 6/13/2025
F2D266B88A494B3...

Slingshot Biosciences Internal Part Number: 11-0618v1