

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102955
Date Issued	APR2026
Expiration Date*	APR2028

***Product has a shelf life of six months from the date of thaw. Follow the Date of Expiration if it occurs before shelf life expiration.**



Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)	
7AAD	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Ghost	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No

Section B: MFI*

Population	MFI Specification	Results (Pass/Fail)	
Negative for 7AAD	≤ 500	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail
Positive for 7AAD	$\geq 20,000$	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail
Negative for Ghost	$\leq 15,000$	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail
Positive for Ghost	$\geq 200,000$	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail

***when analyzed using Cytex Aurora 5 Laser Cytometer under recommended gain settings**

Signed by:  4/30/2026 Senior QA Analyst
 Approved by:  D62A2FF04F7F446...

Quality Assurance (Name, Signature, Date, Title)